


**Aroclor 1221 on ZB5 15m x .32mm x .25um****Column:** Zebtron<sup>™</sup> ZB-5, GC Cap. Column 15 m x 0.32 mm x 0.25  $\mu$ m, Ea**Phase:** 5% Phenyl 95% Dimethylpolysiloxane**Dimensions:** 15 meters x 0.32 mm x 0.25  $\mu$ m**Order No:** 7EM-G002-11**Oven Profile:** 300 $^{\circ}$ C for 12.5 min 300 $^{\circ}$ C for 17.5 min**Carrier Gas:** Constant Flow Helium, 1.7 mL/min**Injection:** Split 15.0000000000:1 0  $\mu$ L @ 250 $^{\circ}$ C**Detection:** Refractive Index (375 $^{\circ}$ C)  
14569**Products used in this application:****ANALYTES:****1** Aroclor 1221