

Caffeine & Metabolites, pH 7.5 - Luna 5µm C18(2)

Column: Luna® 5 µm C18(2) 100 Å, LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-4252-E0

Elution Type: Gradient

Eluent A: 20mM Ammonium acetate, pH 7.5

Eluent B: Acetonitrile

Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	98	2
	2	15	80	20

Flow Rate: 1 mL/min

Col. Temp.: 25 °C

Detection: Refractive Index @ 0.0000000000 (25 °C)

Analyst Note: SecurityGuard™ Guard Cartridge System extends column lifetime.


- SecurityGuard Cartridges, C18 4 x 3.0mm, 10/Pk Part No.: AJ0-4287

- Holder Part No.: KJ0-4282

15662



Products used in this application:



ANALYTES:

1 Xanthine

Retention Time: 4.059 min

2 Theobromine

Retention Time: 8.84 min

3 Theophylline

Retention Time: 10.54 min

4 Caffeine

Retention Time: 13.14 min

