


**EN 141 05: Free and Total Glycerin Analysis in Biodiesel on ZB-5HT****Column:** Zebtron™ ZB-5HT, GC Cap. Column 15 m x 0.32 mm x 0.10 µm, Ea**Phase:** 5% Phenyl 95% Dimethylpolysiloxane**Dimensions:** 15 meters x 0.32 mm x 0.1 µm**Order No:** 7EM-G015-02**Oven Profile:****Carrier Gas:** Constant Flow Helium, 5.79 cm/sec**Injection:** On-Column 0.000000000:1 0 µL @ 50°C**Detection:** Refractive Index (380°C)**Analyst Note:** A 5 m x 0.53 mm Guard Column was connected to the analytical column**Products used in this application:****ANALYTES:**

- 1 Glycerol
- 2 Butanetriol
- 3 Monoolein
- 4 Tricarpin
- 5 Diolein
- 6 Triolein

