

## Lux Chiral Test Mix 5 (AL0-8412)

**Column:** Lux® 3 µm Cellulose-1, LC Column 250 x 4.6 mm, Ea

**Dimensions:** 250 x 4.6 mm ID

**Order No:** 00G-4458-E0

**Elution Type:** Isocratic

**Eluent A:** Hexane

**Eluent B:** IPA

| Gradient Profile: | Step No. | Time (min) | Pct A | Pct B |
|-------------------|----------|------------|-------|-------|
| 1                 | 1        | 0          | 90    | 10    |

**Flow Rate:** 0.5 mL/min


**Col. Temp.:** ambient

**Detection:** Refractive Index @ 0.0000000000 (ambient)

**Analyst Note:** This Check Standard is used to verify the performance of the following Lux chiral column phases: Cellulose 1 and 2. Ordering part number AL0-8412



Products used in this application:



### ANALYTES:

**1** trans-Stilbene oxide

Retention Time: 12.35 min

**2** trans-Stilbene oxide

Retention Time: 21.25 min

