


8270 Check Standard on ZB-50**Column:** Zebron™ ZB-50, GC Cap. Column 30 m x 0.25 mm x 0.25 µm, Ea**Phase:** 50% Phenyl 50% Dimethylpolysiloxane**Dimensions:** 30 meters x 0.25 mm x 0.25 µm**Order No:** 7HG-G004-11**Oven Profile:****Carrier Gas:** Constant Flow Helium, 0.7 mL/min**Injection:** On-Column 0.000000000:1 0 µL @ 43°C**Detection:** Refractive Index (335°C)**Analyst Note:** Analytes are 10 ppm in Acetone

17811

**Products used in this application:****ANALYTES:**

- 1 Pentachlorophenol
- 2 DFTPP
- 3 Benzidine
- 4 DDE
- 5 DDD
- 6 DDT

