


Aroclor 1242: DIN 51527 on ZB-XLB-HT**Column:** Zebron[™] ZB-XLB-HT, GC Cap. Column 30 m x 0.25 mm x 0.25 μ m, Ea**Phase:** Proprietary XLB Phase**Dimensions:** 30 meters x 0.25 mm x 0.25 μ m**Order No:** 7HG-G024-11**Oven Profile:****Carrier Gas:** Constant Flow Helium, 1.5 mL/min**Injection:** Split 2.0000000000:1 0 μ L @ 250°C**Detection:** Refractive Index (350°C)**Analyst Note:** Total concentration of Aroclors was 90 ppm in iso-octane**Products used in this application:**

Aroclor 1242: DIN 51527 on ZB-XLB-HT

ANALYTES:

- 1 TCMX
- 2 BZ# 31
- 3 BZ# 28
- 4 BZ# 52
- 5 BZ# 101
- 6 BZ# 77
- 7 BZ# 123
- 8 BZ# 118
- 9 BZ# 153
- 10 BZ# 114
- 11 BZ# 105
- 12 BZ# 138
- 13 BZ# 126
- 14 BZ# 167
- 15 BZ# 156
- 16 BZ# 180
- 17 BZ# 157
- 18 BZ# 170
- 19 BZ# 169
- 20 BZ# 189
- 21 DCB

