

HPLC Application

ID No.: 18214

Trans Stilbene Oxide on a Lux 5u Cellulose 2 analytical for 50mm comparison

Column: Lux® 5 µm Cellulose-2, LC Column 250 x 4.6 mm, Ea

Dimensions: 250 x 4.6 mm ID

Order No: 00G-4457-E0

Elution Type: Isocratic

Eluent A: Hexane

Eluent B: IPA

Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	90	10

Flow Rate: 0.5 mL/min

Col. Temp.: ambient

Detection: Refractive Index @ 0.0000000000 (ambient)

Analyst Note: SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, Lux Cellulose-2, 4 x 3.0mm, 10/Pk Part No.: AJ0-8366


- Holder Part No.: KJ0-4282



Products used in this application:



18214



ANALYTES:

1 trans-Stilbene oxide

