


8270 Check Std on ZB-1701P**Column:** Zebron[™] ZB-1701, GC Cap. Column 30 m x 0.25 mm x 0.25 μ m, Ea**Phase:** 14% Cyanopropylphenyl 86% Dimethylpolysiloxane**Dimensions:** 30 meters x 0.25 mm x 0.25 μ m**Order No:** 7HG-G006-11**Oven Profile:****Carrier Gas:** Constant Flow Helium, 0.8 mL/min**Injection:** On-Column 0.000000000:1 0 μ L @ 43°C**Detection:** Refractive Index (335°C)**Analyst Note:** Analytes 10 ppm in Acetone

18418

**Products used in this application:****ANALYTES:**

- 1 Pentachlorophenol
- 2 DFTPP
- 3 Benzidine
- 4 DDE
- 5 DDD
- 6 DDT

