


SecurityGuard Comparisons on Lux 250 x 4.6mm With Cartridge

Column: N/A
Dimensions: N/A
Order No: N/A
Elution Type: N/A
Eluent A: Acetonitrile
Eluent B: 20mM NH₄HCO₃
Gradient

Step No.	Time (min)	Pct A	Pct B
1	0	60	40

Profile:
Flow Rate: 0.7 mL/min
Col. Temp.: 25 °C
Detection: Laser Polarimeter @ 0.0000000000 (ambient)
Analyst Note: Analysis performed without SecurityGuard Cartridge (cf. App IDs: 18823 and 18824) using Prodigy Test Mix (PN: AL0-3045)
 19071



ANALYTES:

- 1 impurity
- 2 Chlorpheniramine
- 3 Chlorpheniramine

