


**Sex hormones on a ZB-1ms 30x0.25x0.25 by GCMS****Column:** Zebtron<sup>™</sup> ZB-1MS, GC Cap. Column 30 m x 0.25 mm x 0.25 µm, Ea**Phase:** 100% Dimethylpolysiloxane**Dimensions:** 30 meters x 0.25 mm x 0.25 µm**Order No:** 7HG-G011-11**Oven Profile:****Carrier Gas:** Constant Flow Helium, 1 mL/min**Injection:** Split 10.0000000000:1 0 µL @ 250°C**Detection:** Refractive Index (320°C)**Analyst Note:** Sample Prep: 50µL of 100µg/mL hormone mix in methanol is reacted with 50µL of 2% methoxyamine in pyridine @ 60°C for 30 minutes. Reaction mixture is evaporated to dryness under nitrogen at room temp and reconstituted in BSTFA. The BSTFA mixture is reacted at 70°C for 30 minutes and then injected into GCMS.**Products used in this application:****ANALYTES:**

- 1 Androsterone
- 2 DHEA
- 3 17-alpha-Estradiol
- 4 Estrone
- 5 17-beta-Estradiol
- 6 Testosterone

