


Kinetex 2.6 μ m XB-C18 50 x 4.6 Prodigy Test With SecurityGuard ULTRA**Column:** Kinetex® 2.6 μ m XB-C18 100 Å, LC Column 50 x 4.6 mm, Ea**Dimensions:** 50 x 4.6 mm ID**Order No:** 00B-4496-E0**Elution Type:** Isocratic**Eluent A:** Water**Eluent B:** Acetonitrile

Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	35	65

Flow Rate: 1.85 mL/min**Col. Temp.:** ambient**Detection:** UV-Vis Abs.-Diode Array (PDA) @ (ambient)
19980**Products used in this application:****ANALYTES:****1** Uracil

Retention Time: 0.24 min

2 Acetophenone

Retention Time: 0.39 min

3 Toluene

Retention Time: 0.661 min

4 Naphthalene

Retention Time: 0.793 min

