


EPA Method 548.1: Endothall by GC/MS on ZB-SemiVolatiles**Column:** Zebron™ ZB-SemiVolatiles, GC Cap. Column 30 m x 0.25 mm x 0.25 µm, Ea**Phase:** 5% Phenyl-Arylene 95% Dimethylpolysiloxane (Enviro-Inert)**Dimensions:** 30 meters x 0.25 mm x 0.25 µm**Order No:** 7HG-G027-11**Oven Profile:** 80 Å°C for 5 min to 260 Å°C @ 10 Å°C/min for 10 min**Carrier Gas:** Constant Flow Helium, 1 mL/min**Injection:** Pulsed 0.000000000:1 0 µL @ 200°C**Detection:** Refractive Index (320°C)**Analyst Note:** Pulsed splitless injection @ 30 psi for 0.55 minutes.**Products used in this application:****ANALYTES:**

- 1 Acenaphthene-d10
- 2 Endothall

