

FAMES In Peanut Oil by GC/FID on Zebron ZB-FAME

Column: Zebron ZB-FAME, GC Cap.Column 60m x 0.25mm x 0.2µm, Ea
Phase: Proprietary Pesticides Phase
Dimensions: 60 meters x 0.25 mm x 0.2 µm
Order No: 7KG-G033-10
Oven Profile: 180 Å°C isothermal

Carrier Gas: Constant Flow Helium, 1.2 mL/min
Injection: Split 100.0000000000:1 0 µL @ 240°C
Detection: Refractive Index (240°C)


Analyst Note: Recommended Accessories:
 Recommended Liner: Zebron PLUS Single Taper with Wool, 4 mm ID
 Liner Part No.: AG2-0A11-05 (for Agilent systems)
 Inlet Seal: AG0-8620 (Gold Plated Easy Seal)
 Septum: AG0-4696 (PhenoRed-400)

Sample Preparation:

1. Weigh 0.2 g of sample into a glass test tube
2. Add 3.8 mL of hexane and vortex
3. Condition SPE tube (Strata® Si-1, 2 g/12 mL, Part No.: 8B-S012-KDG) under vacuum with 6 mL of hexane; do not let bed go dry
4. Load sample with light vacuum
5. Elute with 5 mL of hexane:ethyl acetate (87:13)
6. Evaporate under nitrogen at 35 °C
7. Reconstitute with 4 mL of hexane
8. Add 200 µL of 2 M potassium hydroxide/methanol solution and vortex
9. Leave at room temperature for 5 minutes
10. Add 2 mL of deionized water and vortex
11. Leave to separate until upper layer becomes clear
12. Extract upper layer for GC analysis



Products used in this application:



ANALYTES:

- 1 C13:0
- 2 C16:0
- 3 C18:0
- 4 C18:1 cis 9
- 5 C18:2 trans 9,12
- 6 C20:0
- 7 C20:1 cis 11

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



Phenomenex products are available worldwide.

www.phenomenex.com

support@phenomenex.com