

HPLC Application

ID No.: 8330

Distearin in Palm Oil on Phenogel 5u 100A

Column: Phenogel™ 5 µm 100A, LC Column 300 x 7.8 mm, Ea

Dimensions: 300 x 7.8 mm ID

Order No: 00H-0442-K0

Elution Type: Isocratic

Eluent A: THF

Gradient	Step No.	Time (min)	Pct A
Profile:	1	0	100


Flow Rate: 1 mL/min

Col. Temp.: ambient

Detection: Differential refractometer @ 0.0000000000 (ambient)
8330



Products used in this application:



ANALYTES:

- 1 Palm oil
- 2 Distearin
- 3 Palm oil methyl ester
- 4 Monostearin
- 5 Methanol

